

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite Mobile Speed Enforcement System

Customer: Lincolnshire Police **Serial Number:** LW06225
Certificate Number: 6992
Date of Calibration: 27 April 2023 **Next Calibration:** 26 April 2024
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999951 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming Vertical - Pass
 Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	356.6 ft	488.4 ft
Indicated	156.3 ft	356.6 ft	488.7 ft

Speed ± 1 MPH

Actual	27 MPH	36 MPH	-73 MPH	-62 MPH	49 MPH
Indicated	27 MPH	36 MPH	-73 MPH	-62 MPH	49 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK


L Korcz
Technical Signatory

TRUVELO (UK) LTD