CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:	Lincolnshire Police	**Serial Number:**	LW06225
		Certificate Number:	8007
Date of Calibration:	29 April 2024	**Next Calibration:**	28 April 2025
Place of Calibration:	14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG		

Calibration Results:

Checksum		- CC85
Oscillator Frequency		- 19.999966 MHz
Self Test Sequence		- Pass
External Controls		- Pass
Aiming	Vertical	- Pass
	Horizontal	- Pass
Eye Safety		- Pass

Range ± 0.5 ft			
Actual	156.3 ft	356.6 ft	488.4 ft
Indicated	156.3 ft	356.8 ft	488.7 ft

Speed ± 1 MPH					
Actual	23 MPH	-72 MPH	55 MPH	63 MPH	-36 MPH
Indicated	23 MPH	-72 MPH	55 MPH	63 MPH	-36 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

L Korcz
Technical Signatory

TRUVELO (UK) LTD
Tel:020 8847 4400 email:info@truvelouk.com